

Dodatečné informace k zadávacím podmínkám 2

veřejné zakázky

Psychogalvanometr a mobilní klinický elektroencefalograf (EEG)

Additional Information to Tender Conditions No 2

of the public contract

Psychogalvanometer and mobile clinical electroencephalogram (EEG)

Zadavatel poskytuje následující dodatečné informace k zadávacím podmínkám:

1) Technická specifikace se mění takto:

- Psychogalvanometr nemusí být samostatné zařízení per se, ale i rozšíření k EEG.
- Nepožadujeme, aby EEG bylo lékařským zařízením, uchazeč však musí prokázat, že je vhodné pro vědecké účely, tzn. předloží referenci studie publikované v impaktovaném vědeckém časopise s použitím nabízeného zařízení.

2) Nabídky lze podávat do 18. února 2015 do 10 hodin.

The contracting entity supplies the economic operators with the additional information relating to tender conditions:

1) Technical specification has changed as stated below:

- The psychogalvanometer does not have to be a separate device per se it might be extension of the EEG as well.
- We do not require the EEG device to be a medical device. We demand however to have some sort of assurance of its quality for scientific research: a reference to a study published in an impacted scientific journal that uses the offered device.

2) The bids may be submitted no later than 18 February 2015, 10 AM (2015 02 18 10:00).

Brno 10. 2. 2015

Mgr. Rostislav Gnida

manažer veřejných zakázek

Manager of Public Contracts